

TECHNICAL NOTE

Model HI 6100 Series

Update advisory notice for Firmware version 1.21 or higher.

This advisory notice concerns backwards compatibility of Firmware Version 1.21.

Firmware Version 1.21 and above query attached devices for the number of IT Junction Boxes connected for diagnostics. HI 6300 Firmware Versions below 1.0.1.0 do not recognize the query and the number of IT Junction Boxes connected is assumed to be none.

Firmware Version 1.21 adds support for discovering multiple attached devices. HI 6300 Firmware Versions below 1.0.1.0 do not cooperate with the discovery process and cause the discovery process to fail.